

### SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017 & ANSI/NCSL Z540-1-1994

### UHV TECH SERVICES INC. 774 Whittier Hwy South Tamworth, NH 03883 Lisa Crabtree Phone: 603 284 6306

### CALIBRATION

Valid To: October 31, 2025

Certificate Number: 4265.01

-

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

#### I. Fluid Quantities

Parameter/Equipment	Range	CMC <sup>2, 3, 4</sup> (±)	Comments
Leak Rate – Measuring Equipment Helium Gas Leak Standards	(0.1 x 10 <sup>-8</sup> to 1.0 x 10 <sup>-5</sup> ) mbar-l/s	6.0 %	Method: Helium leak standard comparison calibration using a helium mass spectrometer
Leak Test – Measure Test Components	(> 0.1 to 5 x 10 <sup>-12</sup> ) mbar-l/s	6.6 %	Method: Test component connected to helium mass spectrometer under vacuum
	(> 0.1 to 1 x 10 <sup>-7</sup> ) mbar-l/s	6.6 %	Method: Sniffer attached to helium mass spectrometer

<sup>1</sup> This laboratory offers commercial calibration services.

(A2LA Cert. No. 4265.01) 09/13/2023

1\_\_\_\_

Page 1 of 2

- <sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of k = 2. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- <sup>3</sup> In the statement of CMC, percentages are percentages of reading, unless otherwise indicated.
- <sup>4</sup> The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

An





# **Accredited Laboratory**

A2LA has accredited

## **UHV TECH SERVICES INC.**

### South Tamworth, NH

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 13<sup>th</sup> day of September 2023.

Mr. Trace McInturff, Vice President, Accreditation Services For the Accreditation Council Certificate Number 4265.01 Valid to October 31, 2025

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.