



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

UHV TECH SERVICES INC.
774 Whittier Hwy
South Tamworth, NH 03883
Lisa Crabtree Phone: 603 284 6306

CALIBRATION

Valid To: October 31, 2027

Certificate Number: 4265.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with R205 – A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations¹:

I. Fluid Quantities

Parameter/Equipment	Range	CMC ^{2,3,4} (±)	Comments
Leak Rate – Measuring Equipment Helium Gas Leak Standards	(0.1×10^{-8} to 1.0×10^{-5}) mbar-l/s	6.5 %	Method: Helium leak standard comparison calibration using a helium mass spectrometer
Leak Test – Measure Test Components	(> 0.1 to 5×10^{-12}) mbar-l/s (> 0.1 to 1×10^{-7}) mbar-l/s	6.6 % 6.6 %	Method: Test component connected to helium mass spectrometer under vacuum Method: Sniffer attached to helium mass spectrometer

¹ This laboratory offers commercial calibration services.

- ² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- ³ In the statement of CMC, percentages are percentages of reading, unless otherwise indicated.
- ⁴ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



Accredited Laboratory

A2LA has accredited

UHV TECH SERVICES INC.

South Tamworth, NH

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets the requirements of R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system
(refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 23rd day of October 2025.

A blue ink signature of Mr. Trace McInturff.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 4265.01
Valid to October 31, 2027

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.